

Sponsor:

Epoxy Technology
14 Fortune Drive
Billerica, Massachusetts 01821

Date of Test Completion: January 10, 2013

Project Numbers: 12 – 3382
12 – 4004
12 – 5447

Page: 1 of 2

ATTN: Robin Dickie

**Certificate of Compliance
ISO 10993 Biological Tests**

Name EPO-TEK 301-2 Cured 16 hours at 45°C

CAS/Code Number Not Supplied by Sponsor (N/S)

Lot/Batch Number Not Supplied by Sponsor (N/S)

L929 MEM ELUTION CYTOTOXICITY (ISO) –

Toxikon Project 12-3382-G1: The purpose of the MEM Elution is to determine biological reactivity of the mammalian cell culture (L929) in response to the test article extract via microscopic observation. The test article is considered non-cytotoxic and meets the requirements of the MEM Elution Test, ISO 10993-5. *Reference: Biological Evaluation of Medical Devices – Part 5: Tests for In Vitro Cytotoxicity, ISO 10993-5:2009.*

HEMOLYSIS (ASTM) –

Toxikon Project 12-4004-G1: The purpose of this assay is to evaluate the hemolytic potential of the test article. Hemolytic activity of the test article with rabbit blood indicated that the test article was non-hemolytic. (< 5%).

Reference: ASTM F 756-08, Standard Practice for Assessment of Hemolytic Properties of Materials, 2008

KLIGMAN MAXIMIZATION TEST (ISO) –

Toxikon Project 12-5447-G1: The test is used as a procedure for screening of contact allergens in guinea pigs and extrapolating the results to humans. The test article extracts in the USP 0.9% Sodium Chloride for Injection (NaCl) and Cottonseed Oil (CSO) are evaluated for allergenic potential or sensitizing capacity. Based on the scoring and other evaluation standards set by the study protocol, the (NaCl) and (CSO) extracts did not elicit any significant reaction during challenge exposure following an induction phase. The test article is not considered a sensitizer and passes the criteria set in the study protocol.

Reference: Biological Evaluation of Medical Devices – Part 10: Tests for Irritation and Skin Sensitization, ISO 10993-10, 2010.

Project Number: 12 – 3382
12 – 4004
12 – 5447

Page: 2 of 2

**Certificate of Compliance
ISO 10993 Biological Tests**

INTRACUTANEOUS INJECTION (ISO) –

Toxikon Project 12–5447–G2: The purpose of this test is to evaluate the irritation potential of the test article extracts in rabbits after intracutaneous injection. USP 0.9% Sodium Chloride for Injection (NaCl), cottonseed Oil (CSO), did not produce a significantly greater biological reaction than blank extract when injected intracutaneously into rabbits. The test article is not considered an irritant and passes the criteria set forth by the protocol,

Reference: Biological Evaluation of Medical Devices – Part 10: Test for Irritation and Skin Sensitization, ISO 10993–10, 2010

ACUTE SYSTEMIC INJECTION (ISO) –


Toxikon Project 12–5447–G3: The purpose of this assay is to evaluate the test article extracts for potential toxic effects as a result of single dose systemic injection in mice. Test article extracted in USP 0.9% Sodium Chloride for Injection (NaCl), Cottonseed Oil (CSO) did not produce a significantly greater biological reaction than blank extract when injected into mice. The test article did not show greater biological reactivity compared to the control material.

Reference: Biological Evaluation of Medical Devices – Part 11: Tests for Systemic Toxicity, ISO 10993–11:2006

MUSCLE IMPLANTATION TEST (ISO) –

Toxikon Project 12–5447–G4: The purpose of the implant test is to evaluate local toxicity from direct exposure to the test article. The test article is implanted in the paravertebral muscle tissue of New Zealand White rabbits for a period of two weeks. The results indicate that the test article does not demonstrate any remarkable difference as compared to the control implant sites in local tissue responses and the potential to induce local toxic effects.

Reference: Biological Evaluation of Medical Devices – Part 6: Tests for Local Effects After Implantation, ISO 10993–6: 2007.



Director of Biocompatibility Services



Quality Assurance

Date of Certificate: January 18, 2013